

The Radiation Alert[®] Sentry EC is a Personal Alarming Dosimeter / Ratemeter designed to ensure the safety of personnel that work in occupations with potential x-ray or gamma exposure.

The pocket sized unit has an energy compensated tube for a linear response to gamma and built in memory for recording data points for tracking accumulated exposure. The included SentryCom Software enables you to generate incident reconstruction for analysis. The software also permits you to easily set the vibrating and audio alert to your desired levels for dose (>1.0 mR/10 µSv) and dose rate (>1.0 mR/hr/10 µSv/hr).

Use the audio switch to choose between audible clicks with each count recieved or a discreet silent mode.

Detector

Energy Compensated GM Tube **Dose Rate Linearity** Better than ±15% up to 15 R **Energy Response** Down to 20 KeV Accuracy Typically $\pm 15\%$ of reading (Cs¹³⁷) Gamma Sensitivity 1.5 cps/mR/hr (Co⁶⁰⁾

Switch Functions

Power, Vibrate & Audible Alert, Silent Vibrating Only Alert, and Audio Clicks On/Off **Operating Range** Dose Rate: 0.1 - 15 R/hr Accumulated Dose: 0.1 - 65 R Audible Alarms Accumulated Dose & Dose Rate, 90db @ 1 ft. Alarm Thresholds Default: Dose 500 mR/Rate 50 mR/hr Sievert Option: Dose 5000 µSv/Rate 500 µSv/hr Alarm and warning levels are user selectable with the optional SentryCom Software. **Power Requirements**

9-Volt Alkaline (included). Approx. 1500 hrs. at

normal background. Enclosure Alodized Aluminum Housing w/ Metal Belt Clip Outputs Mini-USB



SE INTERNATIONAL INC RADIATION[®] A • L • E • R • T

1-800-293-5759 www.seintl.com

P.O. Box 39, 436 Farm Rd. Summertown, TN 38483 Fax: 931-964-3564 radiationinfo@seintl.com


NTERNATIONAL

THE SENTRY EC RADIATION ALERT ® PERSONAL ALARMING RATE METER & DOSIMETER



The rugged Alodized Aluminum Housing protects the unit in tough industrial environments

Sentry USB Energy Compensated GM Energy Response



SentryCom Software

The optional SentryCom Software enables you to easily integrate The Sentry into your radiation safety program by allowing you to select the thresholds alarms/warnings to your desired levels for dose (>1.0 mR/10 μ Sv) and dose rate (>1.0 mR/hr/10 μ Sv/hr). You can set the Cal Factor, and the data logging frequency, and download the accumulated dose data as a text file. The report includes a delimited time and date stamp, exposure, and power cycle identifiers for easy recording and incident analysis in most spreadsheet programs. Units of measurement can be changed between mR/ hr and μ Sv/hr to fit your application and reporting needs.

The SentryCom Software comes complete with mini-USB cable.

Sentry Com 1.1	
4 Close COM Port	About Lose
Serial Number Firmware Version SENTRY 1.07	
Settings Calibrate Data Log	
Retrieve Current Settings	
Dose Warning Level 0 mR	Rate Warning Level 0 mR/h
Dose Alarm Level 50 mR	Rate Alarm Level mR/h
Sensor Failure if no counts for 5 min	Data Logging Frequency
Units (• mB and mB/h	Load Defaults
Ο μSv and μSv/h	Update With New Settings
Retrieve Date and Time Datalogger Date and time = 2010-10-28 12:13:24 Sentry clock is off by 00:00:00	
Update Date and Time using PC Clock	
Set Date and Time to: 10/28/2010 12:13:21 PM	
Today Now	

P.O. Box 39, 436 Farm Rd. Summertown, TN 38483 1-800-293-5759 | Fax: 931-964-3564 www.seintl.com | radiationinfo@seintl.com